



National Accreditation Board for Testing and Calibration Laboratories

SCOPE OF ACCREDITATION

Laboratory Name :

ELECTRONICS REGIONAL TEST LABORATORY (EAST), BLOCK: DN, SECTOR-V, SALT LAKE CITY, KOLKATA, WEST BENGAL, INDIA

Accreditation Standard

ISO/IEC 17025:2017

Certificate Number

CC-2008

Page No

1 of 59

Validity

21/01/2022 to 20/01/2024

Last Amended on

02/02/2022

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
Permanent Facility					
1	ELECTRO-TECHNICAL-Alternating Current (< 1 GHz) (Measure)	AC Current at 10 Hz - 1 kHz	Using DMM, REFERENCE MULTIMETER (8.5 Digit), AC MEASUREMET STANDARD ,THREE PHASE COMPARATOR by Direct/Comparison Method	10 µA to 1 A	0.05% to 0.016%
2	ELECTRO-TECHNICAL-Alternating Current (< 1 GHz) (Measure)	AC Current at 10 Hz - 5 kHz	Using DMM, REFERENCE MULTIMETER (8.5 Digit), AC MEASUREMET STANDARD, THREE PHASE COMPARATOR by Direct/Comparison Method	1 A to 20 A	0.016 % to 0.035 %
3	ELECTRO-TECHNICAL-Alternating Current (< 1 GHz) (Measure)	AC Current at 50 Hz	Using Three phase comparator by Direct Method	0.01 A to 120 A	0.013%



National Accreditation Board for Testing and Calibration Laboratories

SCOPE OF ACCREDITATION

Laboratory Name : ELECTRONICS REGIONAL TEST LABORATORY (EAST), BLOCK: DN, SECTOR-V, SALT LAKE CITY, KOLKATA, WEST BENGAL, INDIA

Accreditation Standard ISO/IEC 17025:2017

Certificate Number CC-2008

Validity 21/01/2022 to 20/01/2024

Page No 2 of 59

Last Amended on 02/02/2022

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
4	ELECTRO-TECHNICAL-Alternating Current (< 1 GHz) (Measure)	AC Energy 1Ph & 3Ph (50Hz) (active and reactive) , UPF - 0.5 PF (lead & Lag)	Using Three Phase Comparator by Direct / Comparison Method:	60 V to 240 V, 10 mA to 120 A, UPF to 0.5 (lead and Lag)	0.012 % to 0.023 %
5	ELECTRO-TECHNICAL-Alternating Current (< 1 GHz) (Measure)	AC High Voltage at 50 Hz	Using DMM & HV Probe By Direct/ Comparison Method	>1 kV to 28 kV	6%
6	ELECTRO-TECHNICAL-Alternating Current (< 1 GHz) (Measure)	AC Power 1Ph & 3Ph (50Hz) (active and reactive) 60 V to 240V, 10 mA to 120A , UPF - 0.5 PF (lead & Lag)	Using Three Phase Comparator by Direct / Comparison Method:	0.03W to 86.4kW \\(0.03VAR to 86.4kVAR)	0.012% to 0.023%
7	ELECTRO-TECHNICAL-Alternating Current (< 1 GHz) (Measure)	AC Resistance at 1kHz	Using RLC DIGIBRIDGE, Standard AC/DC resistance By Direct/ Comparison Method	1 ohm to 100 kohm	0.3 to 0.1%



National Accreditation Board for Testing and Calibration Laboratories

SCOPE OF ACCREDITATION

Laboratory Name : ELECTRONICS REGIONAL TEST LABORATORY (EAST), BLOCK: DN, SECTOR-V, SALT LAKE CITY, KOLKATA, WEST BENGAL, INDIA

Accreditation Standard ISO/IEC 17025:2017

Certificate Number CC-2008

Validity 21/01/2022 to 20/01/2024

Page No 3 of 59

Last Amended on 02/02/2022

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
8	ELECTRO-TECHNICAL-Alternating Current (< 1 GHz) (Measure)	AC Voltage at 10 Hz -40 Hz	Using IVD, REFERENCE MULTIMETER (8.5 Digit), THERMAL VOLTAGE CONVERTER AC MEASUREMET STANDARD by Direct/ Comparison Method:	1 mV to 1000 V	0.5% to 0.025%
9	ELECTRO-TECHNICAL-Alternating Current (< 1 GHz) (Measure)	AC Voltage at 100kHz-1MHz	Using Multifunction Calibrator & IVD ,AC MEASUREMET STANDARD ,THERMAL VOLTAGE CONVERTER by Direct/Comparison Method	1 mV to 1 V	1.34% to 0.1%
10	ELECTRO-TECHNICAL-Alternating Current (< 1 GHz) (Measure)	AC Voltage at 10kHz-100kHz	Using IVD, REFERENCE MULTIMETER (8.5 Digit), THERMAL VOLTAGE CONVERTER, AC MEASUREMET STANDARD By Direct/ Comparison Method:	1 mV to 1 V	0.43 % to 0.014 %



National Accreditation Board for Testing and Calibration Laboratories

SCOPE OF ACCREDITATION

Laboratory Name :

ELECTRONICS REGIONAL TEST LABORATORY (EAST), BLOCK: DN, SECTOR-V, SALT LAKE CITY, KOLKATA, WEST BENGAL, INDIA

Accreditation Standard

ISO/IEC 17025:2017

Certificate Number

CC-2008

Page No

4 of 59

Validity

21/01/2022 to 20/01/2024

Last Amended on

02/02/2022

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
11	ELECTRO-TECHNICAL-Alternating Current (< 1 GHz) (Measure)	AC Voltage at 10kHz-100kHz	Using IVD, REFERENCE MULTIMETER (8.5 Digit) THERMAL VOLTAGE CONVERTER, AC MEASUREMET STANDARD By Direct/ Comparison Method:	1 V to 100 V	0.014% to 0.016%
12	ELECTRO-TECHNICAL-Alternating Current (< 1 GHz) (Measure)	AC Voltage at 40Hz - 10kHz	Using IVD, REFERENCE MULTIMETER (8.5 Digit), THERMAL VOLTAGE CONVERTER, AC MEASUREMET STANDARD By Direct/ Comparison Method:	1 mV to 1000 V	0.5% to 0.012 %
13	ELECTRO-TECHNICAL-Alternating Current (< 1 GHz) (Measure)	Capacitance at 1 kHz	Using Capacitance Measuring Assembly REFERENCE STD CAPACITOR, RLC DIGIBRIDGE Direct/Comparison	1 pF to 1.0 mF	0.04 % to 1.2 %



National Accreditation Board for Testing and Calibration Laboratories

SCOPE OF ACCREDITATION

Laboratory Name : ELECTRONICS REGIONAL TEST LABORATORY (EAST), BLOCK: DN, SECTOR-V, SALT LAKE CITY, KOLKATA, WEST BENGAL, INDIA

Accreditation Standard ISO/IEC 17025:2017

Certificate Number CC-2008

Validity 21/01/2022 to 20/01/2024

Page No 5 of 59

Last Amended on 02/02/2022

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
14	ELECTRO-TECHNICAL-Alternating Current (< 1 GHz) (Measure)	Capacitance at 1 kHz	Using Capacitance Measuring Assembly Reference STD CAPACITOR ,RLC DIGIBRIDGE Direct/Comparison method	1.0 mF to 10 mF	1.2 % to 0.3 %
15	ELECTRO-TECHNICAL-Alternating Current (< 1 GHz) (Measure)	High Frequency AC Voltage 100MHz -1 GHz	Using RF MILLIVOLTMETER by Direct/Comparison Method	10 mV to 7 V	3.5%
16	ELECTRO-TECHNICAL-Alternating Current (< 1 GHz) (Measure)	High Frequency AC Voltage 1 MHz -100 MHz	Using RF MILLIVOLTMETER by Direct/Comparison Method	10 mV to 10 V	3.3%
17	ELECTRO-TECHNICAL-Alternating Current (< 1 GHz) (Measure)	Inductance at 1 kHz	Using RLC DIGIBRIDGE Reference STD Capacitor by Direct/Comparison Method	100 µH to 10 H	0.2 % to 0.06 %
18	ELECTRO-TECHNICAL-Alternating Current (< 1 GHz) (Measure)	Power Factor/Phase Angle 50Hz At 240V ,5A	Using Three Phase Comparator by Direct/comparison Method	+/- 0.1 - 1.0 (lag & lead) to 0°to 180° (Lead & Lag)	0.01 %



National Accreditation Board for Testing and Calibration Laboratories

SCOPE OF ACCREDITATION

Laboratory Name :

ELECTRONICS REGIONAL TEST LABORATORY (EAST), BLOCK: DN, SECTOR-V, SALT LAKE CITY, KOLKATA, WEST BENGAL, INDIA

Accreditation Standard

ISO/IEC 17025:2017

Certificate Number

CC-2008

Page No

6 of 59

Validity

21/01/2022 to 20/01/2024

Last Amended on

02/02/2022

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
19	ELECTRO-TECHNICAL-Alternating Current (< 1 GHz) (Source)	AC Current 10 Hz - 5 KHz	Using Multifunction Calibrator ,REFERENCE MULTIMETER (8.5 Digit), AC measurement standard & Shunt by Direct/ Comparison Method	10 μ A to 1A	0.08 % to 0.05 %
20	ELECTRO-TECHNICAL-Alternating Current (< 1 GHz) (Source)	AC Current 10 Hz - 5 KHz	Using Multifunction CALIBRATOR , REFERENCE MULTIMETER (8.5 Digit), AC measurement standard & Shunt by Direct/ Comparison Method	1A to 20 A	0.05 % to 0.035 %
21	ELECTRO-TECHNICAL-Alternating Current (< 1 GHz) (Source)	AC Current At 50Hz	Using Three phase comparator by Direct Method	10 mA to 100 A	0.013%
22	ELECTRO-TECHNICAL-Alternating Current (< 1 GHz) (Source)	AC Current by using current coil At 50 Hz	Using Multifunction Calibrator and Current Coil by Direct Method	20 A to 1000 A	0.55 %



National Accreditation Board for Testing and Calibration Laboratories

SCOPE OF ACCREDITATION

Laboratory Name : ELECTRONICS REGIONAL TEST LABORATORY (EAST), BLOCK: DN, SECTOR-V, SALT LAKE CITY, KOLKATA, WEST BENGAL, INDIA

Accreditation Standard ISO/IEC 17025:2017

Certificate Number CC-2008

Validity 21/01/2022 to 20/01/2024

Page No 7 of 59

Last Amended on 02/02/2022

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
23	ELECTRO-TECHNICAL-Alternating Current (< 1 GHz) (Source)	AC Energy 1Ph & 3Ph (50Hz) (active and reactive) , UPF - 0.5 PF (lead & Lag)	Using Three Phase Comparator by Direct / Comparison Method	60 V to 240 V, 10 mA to 120A, UPF to 0.5 (lead & lag) to	0.012 % to 0.023 %
24	ELECTRO-TECHNICAL-Alternating Current (< 1 GHz) (Source)	AC Power 1Ph & 3Ph (50Hz) (active and reactive) , 60 V to 240V ,10 mA to 120A , UPF - 0.5 PF (lead & Lag)	Using Three Phase Comparator by Direct / Comparison Method	(0.03W to 86.4kW) \ (0.03VAR to 86.4kVAR)	0.012% to 0.023%
25	ELECTRO-TECHNICAL-Alternating Current (< 1 GHz) (Source)	AC Resistance at 1 kHz	Using AC/DC RES.STD,LCR Digibridge by Direct / Comparison Method	1 ohm, 10 ohm , 100 ohm,1 kohm, 10 kohm	0.007%
26	ELECTRO-TECHNICAL-Alternating Current (< 1 GHz) (Source)	AC Voltage 10 Hz - 45 Hz	Using Multifunction Calibrator & IVD, AC MEASUREMET STANDARD , THERMAL VOLTAGE CONVERTER by Direct/Comparison Method:	1 mV to 1000 V	0.7% to 0.025%



National Accreditation Board for Testing and Calibration Laboratories

SCOPE OF ACCREDITATION

Laboratory Name : ELECTRONICS REGIONAL TEST LABORATORY (EAST), BLOCK: DN, SECTOR-V, SALT LAKE CITY, KOLKATA, WEST BENGAL, INDIA

Accreditation Standard ISO/IEC 17025:2017

Certificate Number CC-2008

Validity 21/01/2022 to 20/01/2024

Page No 8 of 59

Last Amended on 02/02/2022

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
27	ELECTRO-TECHNICAL- Alternating Current (< 1 GHz) (Source)	AC Voltage 10 kHz - 50 kHz	Using Multifunction Calibrator & IVD AC MEASUREMET STANDARD THERMAL VOLTAGE CONVERTER by Direct/Comparison Method	1 mV to 1 V	0.2% to 0.014%
28	ELECTRO-TECHNICAL- Alternating Current (< 1 GHz) (Source)	AC Voltage 10kHz-50kHz	Using Multifunction Calibrator & IVD AC MEASUREMET STANDARD THERMAL VOLTAGE CONVERTER by Direct/Comparison Method::	1 V to 100 V	0.014% to 0.02%
29	ELECTRO-TECHNICAL- Alternating Current (< 1 GHz) (Source)	AC Voltage 45Hz - 10kHz	Using Multifunction Calibrator & IVD AC MEASUREMET STANDARD THERMAL VOLTAGE CONVERTER by Direct/Comparison Method:	1 mV to 100 V	0.4 % to 0.008 %



National Accreditation Board for Testing and Calibration Laboratories

SCOPE OF ACCREDITATION

Laboratory Name : ELECTRONICS REGIONAL TEST LABORATORY (EAST), BLOCK: DN, SECTOR-V, SALT LAKE CITY, KOLKATA, WEST BENGAL, INDIA

Accreditation Standard ISO/IEC 17025:2017

Certificate Number CC-2008

Validity 21/01/2022 to 20/01/2024

Page No 9 of 59

Last Amended on 02/02/2022

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
30	ELECTRO-TECHNICAL- Alternating Current (< 1 GHz) (Source)	AC Voltage 45Hz - 10kHz	Using Multifunction Calibrator & IVD, AC MEASUREMET STANDARD THERMAL VOLTAGE CONVERTER by Direct/Comparison Method:	100 V to 1000 V	0.008% to 0.02%
31	ELECTRO-TECHNICAL- Alternating Current (< 1 GHz) (Source)	AC Voltage 50kHz -1MHz	Using Multifunction Calibrator & IVD AC MEASUREMET STANDARD THERMAL VOLTAGE CONVERTER by Direct/Comparison Method	1 mV to 100 mV	0.05 % to 1.2 %
32	ELECTRO-TECHNICAL- Alternating Current (< 1 GHz) (Source)	AC Voltage 50kHz -1MHz	Using Multifunction Calibrator & IVD AC MEASUREMET STANDARD THERMAL VOLTAGE CONVERTER by Direct/Comparison Method::	100 mV to 10 V	0.25 % to 0.12 %
33	ELECTRO-TECHNICAL- Alternating Current (< 1 GHz) (Source)	Capacitance 1 kHz	Using REFERENCE STD CAPACITOR, RLC digibridge By Direct Method	1.0 mF, 10 mF	0.05% to 0.38 %



National Accreditation Board for Testing and Calibration Laboratories

SCOPE OF ACCREDITATION

Laboratory Name : ELECTRONICS REGIONAL TEST LABORATORY (EAST), BLOCK: DN, SECTOR-V, SALT LAKE CITY, KOLKATA, WEST BENGAL, INDIA

Accreditation Standard ISO/IEC 17025:2017

Certificate Number CC-2008 **Page No** 10 of 59

Validity 21/01/2022 to 20/01/2024 **Last Amended on** 02/02/2022

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
34	ELECTRO-TECHNICAL- Alternating Current (< 1 GHz) (Source)	Capacitance 1 kHz	Using STD "Capacitor", REFERENCE STD CAPACITOR,RLC Digibridge By Direct / Comparison Method	10 pF to 1000pF	0.003% to 0.015%
35	ELECTRO-TECHNICAL- Alternating Current (< 1 GHz) (Source)	Capacitance 1 kHz	Using REFERENCE STD CAPACITOR, RLC Digibridge By Direct Method	1000 pF to 1 mF	0.015% to 0.05 %
36	ELECTRO-TECHNICAL- Alternating Current (< 1 GHz) (Source)	High Frequency AC Voltage 1 MHz -1 GHz	Using Multifunction Calibrator, RF Voltmeter with Insertion Unit, Signal Generator & RF Amplifier.,1000MHz, RF Attenuator by Direct / Comparison Method	10 mV to 7 V	3.5 %
37	ELECTRO-TECHNICAL- Alternating Current (< 1 GHz) (Source)	Inductance 1 kHz	Using Reference STD INDUCTOR , RLC DIGIBRIDGE by Direct / Comparison Method	100 µH to 100 mH	0.1% to 0.03%
38	ELECTRO-TECHNICAL- Alternating Current (< 1 GHz) (Source)	Inductance 1 kHz	Using Reference STD INDUCTOR , RLC DIGIBRIDGE by Direct / Comparison Method	100 mH to 10H	0.03% to 0.09%



National Accreditation Board for Testing and Calibration Laboratories

SCOPE OF ACCREDITATION

Laboratory Name : ELECTRONICS REGIONAL TEST LABORATORY (EAST), BLOCK: DN, SECTOR-V, SALT LAKE CITY, KOLKATA, WEST BENGAL, INDIA

Accreditation Standard ISO/IEC 17025:2017

Certificate Number CC-2008

Validity 21/01/2022 to 20/01/2024

Page No 11 of 59

Last Amended on 02/02/2022

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
39	ELECTRO-TECHNICAL-DIRECT CURRENT (Measure)	DC Current	Using Current Shunt & Standard Resistance, Current Source & DMM REFERENCE MULTIMETER (8.5 Digit) by Direct/Comparison Method:	10 μ A to 1 mA	0.006 % to 0.002 %
40	ELECTRO-TECHNICAL-DIRECT CURRENT (Measure)	DC Current	Using Current Shunt & Standard Resistance, Current Source & DMM REFERENCE MULTIMETER (8.5 Digit) by Comparison,V/R Method:	20 A to 850 A	0.005 % to 0.05 %
41	ELECTRO-TECHNICAL-DIRECT CURRENT (Measure)	DC Current :	Using Current Shunt & Standard Resistance, Current Source & DMM REFERENCE MULTIMETER (8.5 Digit) by Direct/Comparison Method:	1 mA to 20 A	0.002 % to 0.005 %



National Accreditation Board for Testing and Calibration Laboratories

SCOPE OF ACCREDITATION

Laboratory Name :

ELECTRONICS REGIONAL TEST LABORATORY (EAST), BLOCK: DN, SECTOR-V, SALT LAKE CITY, KOLKATA, WEST BENGAL, INDIA

Accreditation Standard

ISO/IEC 17025:2017

Certificate Number

CC-2008

Page No

12 of 59

Validity

21/01/2022 to 20/01/2024

Last Amended on

02/02/2022

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
42	ELECTRO-TECHNICAL-DIRECT CURRENT (Measure)	DC High Voltage :	Using HV probe and and Digital Multimeter by Direct/Comparison method	>1 kV to 40 kV	2.5 %
43	ELECTRO-TECHNICAL-DIRECT CURRENT (Measure)	DC Resistance	Using DMM, REFERENCE MULTIMETER (8.5 Digit), Standard Resistance Set, High Resistance Meter by V/I, Direct /Comparison ,Method:	0.0001 ohm to 0.001 ohm	0.03 % to 0.004 %
44	ELECTRO-TECHNICAL-DIRECT CURRENT (Measure)	DC Resistance :	Using DMM REFERENCE MULTIMETER (8.5 Digit), Standard Resistance Set, High Resistance Meter by Direct/Comparison , V/I Method:	0.001 ohm to 1 Mohm	0.004 % to 0.001 %
45	ELECTRO-TECHNICAL-DIRECT CURRENT (Measure)	DC Resistance :	Using DMM REFERENCE MULTIMETER (8.5 Digit), Standard High Resistance Meter by Direct/Comparison Method:	1 Mohm to 20 Gohm	0.001 % to 0.20 %



National Accreditation Board for Testing and Calibration Laboratories

SCOPE OF ACCREDITATION

Laboratory Name : ELECTRONICS REGIONAL TEST LABORATORY (EAST), BLOCK: DN, SECTOR-V, SALT LAKE CITY, KOLKATA, WEST BENGAL, INDIA

Accreditation Standard ISO/IEC 17025:2017

Certificate Number CC-2008

Validity 21/01/2022 to 20/01/2024

Page No 13 of 59

Last Amended on 02/02/2022

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(\pm)
46	ELECTRO-TECHNICAL-DIRECT CURRENT (Measure)	DC Resistance :	Using DMM REFERENCE MULTIMETER (8.5 Digit), Standard Resistance Set, High Resistance Meter by Direct/Comparison Method:	20 Gohm to 1 Tohm	0.2% to 2.5 %
47	ELECTRO-TECHNICAL-DIRECT CURRENT (Measure)	DC Voltage	Using DC Reference Standard , Null Detector , Kelvin Verley Divider, REFERENCE MULTIMETER (8.5 Digit) by Direct/ Comparison method	1 mV to 10 V	0.014 % to 0.0004 %
48	ELECTRO-TECHNICAL-DIRECT CURRENT (Measure)	DC Voltage	Using DC Reference Standard , Null Detector , Kelvin Verley Divider, REFERENCE MULTIMETER (8.5 Digit) by Direct/ Comparison method:	10 μ V to 1 mV	2 % to 0.014 %



National Accreditation Board for Testing and Calibration Laboratories

SCOPE OF ACCREDITATION

Laboratory Name :

ELECTRONICS REGIONAL TEST LABORATORY (EAST), BLOCK: DN, SECTOR-V, SALT LAKE CITY, KOLKATA, WEST BENGAL, INDIA

Accreditation Standard

ISO/IEC 17025:2017

Certificate Number

CC-2008

Page No

14 of 59

Validity

21/01/2022 to 20/01/2024

Last Amended on

02/02/2022

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
49	ELECTRO-TECHNICAL-DIRECT CURRENT (Measure)	DC Voltage :	Using DC Reference Standard , Null Detector , Kelvin Verley Divider, REFERENCE MULTIMETER (8.5 Digit) by Direct/ Comparison method:	10 V to 1000 V	0.0004% to 0.0005 %
50	ELECTRO-TECHNICAL-DIRECT CURRENT (Source)	DC Current	Using CALIBRATOR, STD"RESISTANCE", DMM- REFERENCE MULTIMETER (8.5 Digit) STD Shunt, By Direct/ Comparison Method	1 A to 20 A	0.005 % to 0.009 %
51	ELECTRO-TECHNICAL-DIRECT CURRENT (Source)	DC Current	Using CALIBRATOR, STD"RESISTANCE", REFERENCE MULTIMETER STD Shunt, By Direct/ V/R Method	10 μ A to 1 A	0.01 % to 0.005 %
52	ELECTRO-TECHNICAL-DIRECT CURRENT (Source)	DC Current	Using CALIBRATOR along with current coil By Direct Method	20 A to 1000 A	0.53%



National Accreditation Board for Testing and Calibration Laboratories

SCOPE OF ACCREDITATION

Laboratory Name :

ELECTRONICS REGIONAL TEST LABORATORY (EAST), BLOCK: DN, SECTOR-V, SALT LAKE CITY, KOLKATA, WEST BENGAL, INDIA

Accreditation Standard

ISO/IEC 17025:2017

Certificate Number

CC-2008

Page No

15 of 59

Validity

21/01/2022 to 20/01/2024

Last Amended on

02/02/2022

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
53	ELECTRO-TECHNICAL-DIRECT CURRENT (Source)	DC Current	Using CALIBRATOR along with Transconductance Amplifier, STD Resistance, Reference Multimeter, SYSTEM DC POWER SUPPLY, STD Shunt, By Direct/V/R Method	20 A to 850 A	0.009 % to 0.06 %
54	ELECTRO-TECHNICAL-DIRECT CURRENT (Source)	DC Resistance	Using STD "Resistance" By Direct Method	0.0001 ohm to 100 kohm	0.04 % to 0.001 %
55	ELECTRO-TECHNICAL-DIRECT CURRENT (Source)	DC Resistance	Using STD "Resistance" By Direct Method	10 Mohm to 1 Tohm	0.003 % to 2 %
56	ELECTRO-TECHNICAL-DIRECT CURRENT (Source)	DC Resistance	Using STD "Resistance", By Direct Method	100 kohm to 10 Mohm	0.001 % to 0.003 %



National Accreditation Board for Testing and Calibration Laboratories

SCOPE OF ACCREDITATION

Laboratory Name : ELECTRONICS REGIONAL TEST LABORATORY (EAST), BLOCK: DN, SECTOR-V, SALT LAKE CITY, KOLKATA, WEST BENGAL, INDIA

Accreditation Standard ISO/IEC 17025:2017

Certificate Number CC-2008

Validity 21/01/2022 to 20/01/2024

Page No 16 of 59

Last Amended on 02/02/2022

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
57	ELECTRO-TECHNICAL-DIRECT CURRENT (Source)	DC Voltage	Using MULTIFUNCTION CALIBRATOR, DMM-REFERENCE MULTIMETER (8.5 Digit), DC Voltage Ref STD NULL DETECT, KELVIN VERLEY DIVIDER By Direct / Comparison Method	10 μ V to 10 V	2 % to 0.0003 %
58	ELECTRO-TECHNICAL-DIRECT CURRENT (Source)	DC Voltage	Using MULTIFUNCTION CALIBRATOR, DMM-REFERENCE MULTIMETER (8.5 Digit), DC Voltage Ref STD NULL DETECTOR, KELVIN VERLEY DIVIDER By Direct / Comparison Method	10 V to 1000 V	0.0003 % to 0.0004 %
59	ELECTRO-TECHNICAL-RF/MICROWAVE (1 GHZ AND ABOVE) (Measure)	Attenuation 50MHz-1GHz	Using RF Level Meter, RF MILLIVOLTMETER by Direct/Comparison Method	1 dB to 10 dB	0.15dB



National Accreditation Board for Testing and Calibration Laboratories

SCOPE OF ACCREDITATION

Laboratory Name : ELECTRONICS REGIONAL TEST LABORATORY (EAST), BLOCK: DN, SECTOR-V, SALT LAKE CITY, KOLKATA, WEST BENGAL, INDIA

Accreditation Standard ISO/IEC 17025:2017

Certificate Number CC-2008

Validity 21/01/2022 to 20/01/2024

Page No 17 of 59

Last Amended on 02/02/2022

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
60	ELECTRO-TECHNICAL-RF/MICROWAVE (1 GHZ AND ABOVE) (Measure)	Attenuation 50MHz-1GHz	Using RF Level Meter, RF Millivoltmeter by Direct/Comparison Method	10 dB to 60 dB	0.17dB
61	ELECTRO-TECHNICAL-RF/MICROWAVE (1 GHZ AND ABOVE) (Measure)	RF Power 50MHz - 2GHz	Using RF MILLIVOLTMETER, RF STEP ATTENUATOR Power Sensors, RF Attenuator , RF Amplifier By Direct/ Comparison Method	1 mW to 40 mW	4%
62	ELECTRO-TECHNICAL-RF/MICROWAVE (1 GHZ AND ABOVE) (Measure)	RF Power 50MHz - 2GHz	Using RF MILLIVOLTMETER, RF STEP ATTENUATOR Power Sensors, RF Attenuator , RF Amplifier By Direct/ Comparison Method	1 nW to 1 mW	6%
63	ELECTRO-TECHNICAL-RF/MICROWAVE (1 GHZ AND ABOVE) (Measure)	RF Power 50MHz to 1GHz	Using RF MILLIVOLTMETER, RF STEP ATTENUATOR Power Sensors, RF Attenuator , RF Amplifier By Direct/ Comparison Method	40 mW to 30 W	4%



National Accreditation Board for Testing and Calibration Laboratories

SCOPE OF ACCREDITATION

Laboratory Name : ELECTRONICS REGIONAL TEST LABORATORY (EAST), BLOCK: DN, SECTOR-V, SALT LAKE CITY, KOLKATA, WEST BENGAL, INDIA

Accreditation Standard ISO/IEC 17025:2017

Certificate Number CC-2008 **Page No** 18 of 59

Validity 21/01/2022 to 20/01/2024 **Last Amended on** 02/02/2022

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
64	ELECTRO-TECHNICAL-RF/MICROWAVE (1 GHZ AND ABOVE) (Measure)	RF Power 50MHz-1GHz	Using RF MILLIVOLTMETER RF Attenuator, Amplifier Research RF Amplifier & Signal Gen by Direct / Comparison Method	40 mW to 30 W	4%
65	ELECTRO-TECHNICAL-RF/MICROWAVE (1 GHZ AND ABOVE) (Measure)	VSWR : 50MHz-2GHz	Using SWR BRIDGE WITH OPEN, SHORT & COAXIAL MISMATCH RF MILLIVOLTMETER, By Return Loss Method	1.05 to 3	0.15 to 0.3 in VSWR
66	ELECTRO-TECHNICAL-RF/MICROWAVE (1 GHZ AND ABOVE) (Source)	Attenuation 50MHz - 1GHz	Using RF MILLIVOLTMETER ,Level Meter with Sensor, Power Meter, RF Step Attenuator, RSP By Power Ratio Method	1 dB to 10 dB	0.2 dB
67	ELECTRO-TECHNICAL-RF/MICROWAVE (1 GHZ AND ABOVE) (Source)	Attenuation 50MHz - 1GHz	Using RF Millivoltmeter ,Level Meter with Sensor, Power Meter, RF Step Attenuator by Power Ratio Method	10 dB to 60 dB	0.2dB



National Accreditation Board for Testing and Calibration Laboratories

SCOPE OF ACCREDITATION

Laboratory Name : ELECTRONICS REGIONAL TEST LABORATORY (EAST), BLOCK: DN, SECTOR-V, SALT LAKE CITY, KOLKATA, WEST BENGAL, INDIA

Accreditation Standard ISO/IEC 17025:2017

Certificate Number CC-2008

Validity 21/01/2022 to 20/01/2024

Page No 19 of 59

Last Amended on 02/02/2022

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
68	ELECTRO-TECHNICAL-RF/MICROWAVE (1 GHZ AND ABOVE) (Source)	RF Power 30MHz-2GHz	Using RF MILLIVOLTMETER, RF Attenuator, Amplifier Research RF Amplifier & Signal Gen by Direct / Comparison Method	1 mW to 40 mW	4%
69	ELECTRO-TECHNICAL-RF/MICROWAVE (1 GHZ AND ABOVE) (Source)	RF Power 30MHz-2GHz	Using RF MILLIVOLTMETER RF Attenuator, RF Amplifier & Signal Gen by Direct / Comparison Method	1 nW to 1 mW	6%
70	ELECTRO-TECHNICAL-RF/MICROWAVE (1 GHZ AND ABOVE) (Source)	VSWR 50MHz-1GHz	Using SWR BRIDGE WITH OPEN, SHORT & COAXIAL MISMATCH RF MILLIVOLTMETER, By Return Loss Method	1.05 to 3	0.15 to 0.3 in VSWR
71	ELECTRO-TECHNICAL-TIME & FREQUENCY (Measure)	Frequency	Using frequency Counter GPS Controlled Rubidium Frequency standard Timer/ Counter/ Analyzer,GPS Controlled frequency Standard by Direct / Comparison Method	10 Hz to 20 GHz	5X10-6 to 6x10-10



National Accreditation Board for Testing and Calibration Laboratories

SCOPE OF ACCREDITATION

Laboratory Name :

ELECTRONICS REGIONAL TEST LABORATORY (EAST), BLOCK: DN, SECTOR-V, SALT LAKE CITY, KOLKATA, WEST BENGAL, INDIA

Accreditation Standard

ISO/IEC 17025:2017

Certificate Number

CC-2008

Page No

20 of 59

Validity

21/01/2022 to 20/01/2024

Last Amended on

02/02/2022

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
72	ELECTRO-TECHNICAL-TIME & FREQUENCY (Measure)	Time Interval / Time Period	Using Freq. Counter by Direct/Comparison Method	20 nsec to 2000 sec	0.0002%
73	ELECTRO-TECHNICAL-TIME & FREQUENCY (Source)	Frequency	Using HIGH STABILITY FREQUENCY COUNTER , GPS controlled Frequency Standard, FUNCTION WAVEFORM GENERATOR, Signal Generator by Direct / Comparison Method	10 Hz to 20 GHz	5X10-6 to 6x10-10
74	ELECTRO-TECHNICAL-TIME & FREQUENCY (Source)	Time Interval/Time Period	Using Func. Gen. Freq. Counter by Direct Method	20 nsec to 2000 sec	0.0002%
75	MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.)	Bevel Protractor L.C.: 5 min	Using Angle Gauge Set By comparison method	0 to 360 °	1 Min of Arc